Product name:	Product code:
Medical Mask Bacterial Filtration and Differential Pressure Tester	CMI-IND-0295

Description

Tester for breathability and filtration efficiency of medical face masks, with a sample holder and pressure sensors. Used by mask manufacturers and test laboratories for release and type tests.

Technical specification

- Holder
- Standards

A mask sample is clamped in a holder and air passes through at set flow. Pressure drop is read by sensors. Combined with a particulate filtration test.

Features

NAUGRA®
Civil Laboratory Equipment Manufacturers and Suppliers India Civil Lab Exportrs India
Email us at: export@civilmechanicalindia.com

- Flow 8 L/min
- Pressure 0 to 500 Pa
- Sample 25 mm
- Digital
- Clamp

Technical Specifications

Test Standards	EN 14683, ASTM F2100, ASTM F2101
Flow	8 L/min
Pressure	0 to 500 Pa
Accuracy	+/- 1 Pa
Sample Area	4.9 cm2
Filter	Particulate
Display	LCD
Power	230 V AC, 50 Hz

Standard Accessories