Product name:	Product code:
Sterile Barrier Package Seal Strength Tester for Medical Pouches	CMI-IND-0297

Description

Peel tester for sterilization pouches and sterile barrier packaging seals that records peak and mean peel force. Used by medical device packagers to verify that seals hold after sterilization.

Technical specification

- Peel fixture
- Cutter

Strips of the seal are peeled at set speed and forces recorded. Ensures the sterile barrier holds up. Pull and angle fixtures are available.

Features

NAUGRA®
Civil Laboratory Equipment Manufacturers and Suppliers India Civil Lab Exportrs India
Email us at: export@civilmechanicalindia.com

- Load cell 50 N
- Peel fixture
- Speed 200 mm/min
- Peak and mean
- PC software

Technical Specifications

Test Standards	ASTM F88, ISO 11607-1, EN 868
Load	50 N
Speed	200 mm/min
Width	15 mm
Accuracy	+/- 0.5 %
Stroke	500 mm
Display	LCD
Power	230 V AC, 50 Hz

Standard Accessories