Product name:	Product code:
ACP Fire Core Combustibility Test Apparatus	CMI-IND-0483

Description

Small flame test chamber that applies a burner flame to the panel edge for 15 or 30 seconds and records flame spread on polyethylene and fire retardant core panels. Used by ACP makers and fire test laboratories.

Technical specification

- Burner
- Holder
- Gas gauge

A burner flame is applied to the panel edge for 30 seconds and flame spread is measured. Fire retardant grades are compared with standard PE core panels.

NAUGRA®
Civil Laboratory Equipment Manufacturers and Suppliers India Civil Lab Exportrs India
Email us at: export@civilmechanicalindia.com

Features

- Burner with timer
- Specimen holder vertical
- Stainless chamber
- Exhaust
- Flame height gauge

Technical Specifications

Flame height	20 mm
Application	15 or 30 s
Specimen	90 x 250 mm
Chamber	SS 304
Gas	Propane
Timer	0.1 s
Standards	EN ISO 11925-2, IS 12777 fire test
Power	230 V AC, 50 Hz

Standard Accessories