
Product Code : TCE-0013



Description :

Pat test, brass ring mould

Technical Specification :

SOUNDNESS OF HYDRATED LIME AND GYPSUM PLASTERS: Utilized for the determination of the soundness of hydrated lime, gypsum and building plasters. Consisting of a brass ring mould, 100 mm. diameter by 5 mm deep.

The mould has an inside taper of 5°. Supplied complete with glass base plate

Civil Mechanical India

Website: www.civilmechanicalindia.com, Email: export@civilmechanicalindia.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India, Phone: +91-0171-2643080, +91-0171-2601773