Product Name:

Water-Water Heat Pump Test Rig

Product Code:

EEZ0030



Description:

Water-Water Heat Pump Test Rig

Technical Specification:

Water-Water Heat Pump Test Rig Manufacturer

Range Of Experiments

- To observe the working principal of heat pump cycle.
- To determine Actual cooling C.O.P., Carnot C.O.P., Theoretical C.O.P.,
- Actual heating C.O.P., Relative C.O.P., Ton of Refrigeration Etc.
- To determine the efficiency of the heat pump system.
- Observe the working of refrigeration cycle also.

Experimental Setup:

- Compressor: ½ Tone, Hermetically sealed compressor, Make Emmelsion Coplend/equivalent.
- Condenser: Coiled Tube Heat Exchanger acts as a condenser. Copper coil is merged in water tank to cool down the gas temp. Continuous water flow circulation is required to extract the heat of gases.
- Expansion device : Thermostatic expansion valve.
- Evaporator coil: Coiled Tube Heat Exchanger copper coil is merged in water tank to extract the heat from water tank. Same can be mentioned by electric heater.
- Electric Heater: For Balancing refrigerating effect produced.
- Rotameter: For measuring liquid flow of Refrigerant (Eureka Make).
- Energy Meter: One each for power measurement of compressor & heater.
- Pressure Gauge: 2 Nos. for H.P., L.P. measurement
- HP/LP cutout: 1 No. to suit compressor, Ranko/Equiv.
- Service valve needle type: 1 No.
- Thermostat : 1 No.
- Suitable filter/drier Make Indfoss.

Control Panel:

- 12 Channel Digital Temperature Indicator
- Dimmerstat: 230 V. AC single phase for heating control
- Main On/Off Switch.

• Voltmeter : 0 − 250 V.

• Ammeter : 0 – 15 A.

Civil Mechanical India

Website: www.civilmechanicalindia.com, Email: export@civilmechanicalindia.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India, Phone: +91-0171-2643080, +91-0171-2601773